

**Jurnal Info Kesehatan**

Vol. 22, No. 3, September 2024, pp. 481-493

P-ISSN 0216-504X, E-ISSN 2620-536X

DOI: [10.31965/infokes.Vol22.Iss3.1353](https://doi.org/10.31965/infokes.Vol22.Iss3.1353)Journal homepage: <https://jurnal.poltekkeskupang.ac.id/index.php/infokes>**RESEARCH****Open Access****Effect of The Temperature on The Size of Inhibition Zone of the Clindamycin, Levofloxacin, Tetracycline, and Trimethoprim Activity Against *Staphylococcus aureus* ATCC 25923****Sistiyono<sup>1a</sup>, Marta Atik Martsiningsih<sup>1b</sup>, Menik Kasiyati<sup>1c</sup>, Muji Rahayu<sup>1d</sup>, Rita Rena Pudyastuti<sup>1e</sup>, Suyana<sup>1f</sup>, Ulyy Rahmawati<sup>1g</sup>, Rahmad<sup>2h</sup>, Zulfikar Husni Faruq<sup>1i</sup>, Budi Setiawan<sup>1,3j</sup>**<sup>1</sup> Department of Medical Laboratory Technology, Poltekkes Kemenkes Yogyakarta, Yogyakarta, Indonesia<sup>2</sup> UPTD Health Laboratory of East Kalimantan Province, Samarinda, East Kalimantan, Indonesia<sup>3</sup> Center of Excellent of Innovation Applied Technology on Public Health, Poltekkes Kemenkes Yogyakarta, Yogyakarta, Indonesia<sup>a</sup> Email address: [sistiyono@gmail.com](mailto:sistiyono@gmail.com)<sup>b</sup> Email address: [atikmartsiningsih@gmail.com](mailto:atikmartsiningsih@gmail.com)<sup>c</sup> Email address: [rifani.2010@gmail.com](mailto:rifani.2010@gmail.com)<sup>d</sup> Email address: [hayuningpuji@gmail.com](mailto:hayuningpuji@gmail.com)<sup>e</sup> Email address: [ritapudyastuti1968@gmail.com](mailto:ritapudyastuti1968@gmail.com)<sup>f</sup> Email address: [suryaniajeng.2014@gmail.com](mailto:suryaniajeng.2014@gmail.com)<sup>g</sup> Email address: [ullya.rahmawati@poltekkesjogja.ac.id](mailto:ullya.rahmawati@poltekkesjogja.ac.id)<sup>h</sup> Email address: [sky\\_lights43@yahoo.com](mailto:sky_lights43@yahoo.com)<sup>i</sup> Email address: [faruq@poltekkesjogja.ac.id](mailto:faruq@poltekkesjogja.ac.id)<sup>j</sup> Email address: [budi.setiawan@poltekkesjogja.ac.id](mailto:budi.setiawan@poltekkesjogja.ac.id)

Received: 18 October 2023

Revised: 20 May 2024

Accepted: 12 September 2024

**Abstract**

Antibiotic sensitivity should be tested. In the sensitivity test, there are technical factors that influence the formation of the inhibition zone diameter. Based on several research one of the technical factors that affect the diameter of the inhibition in the disc diffusion method is the temperature incubation of the media, this must be examined so that it can be controlled to ensure the validity of the sensitivity test results. This study aims to determine the mean, difference, and analyze the diameter of the inhibition zone of the antibiotics namely Clindamycin, Levofloxacin, Tetracycline, and Trimethoprim against *Staphylococcus aureus* on Mueller-Hinton agar media with incubation temperatures of 33°C, 34°C, 35°C, 36°C and 37°C for 18 hours. This research is observational, with a cross-sectional design. The data used are primary data with 100 data on the diameter of the antibiotic inhibition zone, obtained by measuring the diameter of the inhibition zone with different incubation temperatures. The selection of antibiotics is based on the mechanism of action of antibiotics inhibiting bacteria namely, the cell wall or membranes that surrounds the bacterial cell; the machineries that make the nucleic acids DNA and RNA and the machinery that produce proteins (the ribosome and associated proteins) with a range of inhibition zones based on Internal Quality Control CLSI. The data will be processed univariately and the Repeated Measure statistical test to determine the significance of the difference in the diameter of the formed inhibition zone using the ANOVA test. The results of the measurement of the inhibition zone diameter on the incubation temperature variation showed a significant difference with p-value 0.000 for Levofloxacin, Tetracycline and Trimethoprim, while for p-value Clindamycin is 0.010. Levofloxacin, Tetracycline, and Trimethoprim antibiotics, the higher the incubation temperature, the average diameter of the inhibition zone is smaller, while for Clindamycin the higher the incubation temperature, the higher the average diameter of the inhibition zone is the same. There is an effect of incubation temperature volume on the diameter of the antibiotic inhibition zone in the disc diffusion method antibiotic sensitivity test. The research indicates that incubation temperature affects the diameter of the antibiotic inhibition zone in disc diffusion tests, underscoring the need for standardized and precise testing conditions to ensure accurate and reliable antibiotic sensitivity results.

**Keywords:** Incubation Temperature, Inhibition Zone, Sensitivity Test, Disk Diffusion.**\*Corresponding Author:**

Budi Setiawan

Department of Medical Laboratory Technology, Poltekkes Kemenkes Yogyakarta, Yogyakarta, Indonesia

Email: [budi.setiawan@poltekkesjogja.ac.id](mailto:budi.setiawan@poltekkesjogja.ac.id)

©The Author(s) 2024. This article is distributed under the terms of the Creative Commons Attribution 4.0 International License (<http://creativecommons.org/licenses/by/4.0/>), which permits unrestricted use, distribution, and reproduction in any medium, provided you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated.

## 1. INTRODUCTION

The main health problems that occur in developing countries are infectious diseases. Microorganisms that cause disease in humans are called pathogenic microorganisms, one of which is pathogenic bacteria (Novard et al., 2019). *Staphylococcus aureus* is a Gram-positive commensal bacterium that is an opportunistic pathogen. These bacteria live in about 30% of the healthy adult population and are mostly found on the skin and mucous membranes. Although *Staphylococcus aureus* is a commensal bacterium, it has the potential to cause various diseases with varying degrees of severity (Jenul & Horswill, 2019).

Bacterial infections can be treated using antibiotics that are bactericidal or bacteriostatic. To treat patients appropriately and adequately, data on sensitivity tests of bacteria that cause infection are needed for various antibiotics available on the market (Nadjamuddin, 2023). Antibiotic sensitivity test using disc diffusion method for *Staphylococcus* sp bacteria can use Mueller-Hinton media with a turbidity standard of 0.5 McFarland. Bacterial growth on the media will be optimal at a temperature of 35°C with ambient air for 16-18 hours of incubation (Zimmer & et all, 2020).

According to several research results on sensitivity testing, technical factors that affect the diameter or size of the inhibition in the disc diffusion method are inoculum density, disc installation time, incubation temperature, incubation time, plate size, and thickness of agar medium and media composition (Lenggu et al., 2020). Research on the effect of incubation temperature on the inhibition zone has also been carried out on bacteria originating from water, which was carried out by incubation at various temperatures and times, including at a temperature of 35°C for 16 hours, 28°C, and 22°C for 24 hours, 22°C. C for 48 hours, and <19°C for >96 hours. The disc method inhibition zone data precision in this study decreased significantly as the incubation temperature decreased and the time increased (Smith et al., 2018).

The selection of antibiotics in this study was based on the function of the mechanism of action of antibiotics in killing or inhibiting the growth of *Staphylococcus aureus* bacteria including, Clindamycin and Tetracycline which work by inhibiting protein synthesis, Levofloxacin which works by inhibiting nucleic acid synthesis, and Trimethoprim which acts as an antimetabolite. In addition, the selection of these four antibiotics was also based on their use as an Internal Quality Control strain of pure bacteria *Staphylococcus aureus* ATCC 25923 (Murray et al., 2021), (Zimmer & et all, 2020).

The main target of laboratory personnel in the field of clinical microbiology is the rapid and accurate identification of pathogenic microorganisms from clinical specimens and providing an accurate interpretation of the results of the sensitivity test of pathogenic bacteria. The accuracy of antibiotic sensitivity test results will help the healing process of patients infected with bacterial diseases (Willey et al., 2017). For this reason, laboratory personnel must continue to carry out internal quality stabilization, one of which is equipment maintenance to ensure the quality of the inspection results.

The microbiological incubator is one of the important instruments in the process of culturing microorganisms or testing for antibiotic sensitivity. This instrument must be periodically tested for performance and calibration so that the quality of the results of laboratory examinations is maintained. However, in reality, there are still laboratory workers who do not perform routine maintenance of microbiological equipment including incubators. Incubators that are not maintained properly will cause high-temperature deviations. the threshold value for electrical safety, the allowable deviation value at the performance output of the maintenance incubator is  $\pm 1$  C. The easiest way to monitor the incubator temperature is to compare the temperature that appears on the incubator screen with a calibrated thermometer, if the


temperature deviation exceeds 2°C, it is necessary to reset the incubator or calibration (Siregar et al., 2018).

The temperature that is not optimal will affect the formation of bacterial inhibition zones in the bacterial sensitivity test (Nadjamuddin, 2023). Unstable mains voltage, unmonitored equipment performance, frequent opening and closing of the incubator door, the incubator filled with media can cause inappropriate temperature output. The importance of the incubation temperature variable in the antibiotic sensitivity test is the basis for conducting research on the Effect of Incubation Temperature on the Diameter of the Inhibitory Zone in the Antibiotic Sensitivity Test of *Staphylococcus aureus*.

This study aims to determine the mean, difference and analyze the difference in the diameter of the inhibition zone of the antibiotics Clindamycin, Levofloxacin, Tetracycline, and Trimethoprim against *Staphylococcus aureus* on Mueller-Hinton agar media with incubation temperatures of 33°C, 34°C, 35°C, 36°C and 37°C for 18 hours.

## 2. RESEARCH METHOD

The type of research used in this research is observational research with a cross-sectional study. This type of research is a type of research conducted without intervention on the research subject (Notoatmodjo, 2005). This study was conducted to observe whether there was a difference in the diameter of the antibiotic inhibition zone against *Staphylococcus aureus* bacteria incubated at different temperatures. The data collection technique used in this study was to measure the diameter of the inhibition zone in the disc diffusion method of antibiotic sensitivity test. Purity test of *S. aureus* bacteria on blood agar plate media can be seen in figure 1 below.



**Figure 1.** Purity Test of *S. aureus* Bacteria on Blood Agar Plate Media

A pure bacterial strain of *Staphylococcus aureus* ATCC 25923 was obtained from the microbiology laboratory at Health Laboratory East Kalimantan Province. The bacterial strain of *Staphylococcus aureus* ATCC 25923 was enriched for 24 hours in Brain Heart Infusion (BHI) media. The cloudy BHI media indicated bacterial growth, then inoculated on an inclined.

Nutrient Agar medium for antibiotic sensitivity testing, and inoculated on two Blood Agar Plate media to identify the purity of the bacteria. All inoculated media were incubated at

35°C for 24 hours. Identification of *Staphylococcus aureus* bacteria was carried out using an automatic device brand BD Phoenix 50i by reacting a 0.5 McFarland bacterial suspension to the PMIC panel substrate reagent.

The technique of collecting samples with Consecutive sampling is done by selecting and making a standard of bacterial turbidity (0.5 McFarland). The diameter of the inhibition zone on the antibiotics Clindamycin, Levofloxacin, Tetracycline, and Trimethoprim was measured against the pure bacterial strain of *Staphylococcus aureus* ATCC 25923 which was formed after incubation at 33°C, 34°C, 35°C, 36°C, and 37°C. The antibiotic sensitivity test was repeated 5 times, so that the number of plates examined for 5 variations in temperature was 25 times. In one plate can be filled with 4 types of antibiotics so that with 5 repetitions can produce 100 data on the diameter of the antibiotic inhibition zone.

The instrument validity test is carried out by means of periodic maintenance and calibration of the equipment as a form of strengthening the internal quality of the laboratory. One of the tools used in this research is an incubator, this tool is well maintained and has been calibrated regularly every year at the microbiology Laboratory, Health Laboratory of East Kalimantan Province by the Calibration Testing Agency. For other tools such as the Nephelometer, the validity of the instrument can also be done by testing the performance of the tool against the turbidity standard available on the brand of the tool. Other validity tests that were also carried out were the Mueller-Hinton media sterility test and the *Staphylococcus aureus* bacteria purity test. Mueller-Hinton media were prepared according to standard procedures, taken as much as 5% and incubated without inoculation at 35°C for 18-24 hours. After 24 hours of incubation, the colony growth was observed on Mueller-Hinton media. If there is no bacterial growth then all media can be continued for antibiotic sensitivity testing, but if there is bacterial growth then all media in one batch cannot be continued for testing.


Data were collected by measuring the diameter of the inhibition zone on the antibiotic sensitivity test that had been determined using the disc diffusion method after being incubated at various temperatures. The data collected were analyzed descriptively and statistical analysis using the Repeated Measures ANOVA test with a degree of error ( $\alpha$ ) of 5%.

### 3. RESULTS AND DISCUSSION

This study used a pure strain of *Staphylococcus aureus* bacteria. The antibiotics used were Clindamycin 2µg, Levofloxacin 5µg, Tetracycline 30µg, and Trimethoprim 5µg. The results of the study in the form of the diameter of the antibiotic inhibition zone against *Staphylococcus aureus* bacteria were analyzed according to the research objectives, namely, to see the mean and difference in the mean of each independent variable, and to see the significance of the difference in the diameter of the inhibition zone formed after incubation at 33°C, 34°C, 35 °C, 36°C, and 37°C.

**Table 1.** Average Diameter of Antibiotic Inhibitory Zone against *Staphylococcus aureus* at Variation of Incubation Temperature

Antibiotic	Average Diameter of Antibiotic Inhibitory Zone (mm)				
	33°C	34°C	35°C	36°C	37°C
Clindamycin	23.33	25.71	24.28	23.85	23.43
Levofloxacin	28.29	27.41	25.87	26.91	25.30
Tetracycline	28.13	27.58	228.80	25.01	25.23
Trimethoprim	25.77	25.59	23.40	23.26	23.32



Noted: a=33°C; b=34°C; c=35°C; d=36°C; e=37°C

**Figure 2.** Inhibition Zone at Various Incubation Temperatures.

**Table 2.** The Difference in Mean Diameter of Antibiotic Inhibitory Zone at Variation of Incubation Temperature to Optimal Temperature

Antibiotic	Average Inhibitory Zone Diameter at Variation of Incubation Temperature		
	Temperature Variation	Optimal Temperature	Deviation
Clindamycin	33°C	35°C	0.95
	34°C		1.43
	36°C		0.43
	37°C		0.85
Levofloxacin	33°C	35°C	2.42
	34°C		1.54
	36°C		1.04
	37°C		0.57
Tetracycline	33°C	35°C	0.67
	34°C		1.22
	36°C		3.79
	37°C		3.57
Trimethoprim	33°C	35°C	2.37
	34°C		219
	36°C		0.14
	37°C		0.08

Table 1 shows the mean diameter of the different inhibition zones of the four types of antibiotics. It can be seen that for Levofloxacin, Tetracycline, and Trimethoprim antibiotics, the higher the incubation temperature, the average diameter of the inhibition zone is smaller, while for Clindamycin the higher the incubation temperature, the higher the average diameter

of the inhibition zone is the same, macroscopically this can be seen in Figure 2. Table 2 shows the difference in the mean diameter of the inhibition zones of each antibiotic. The smallest mean difference is 0.08 (mm) and the largest is 3.79 (mm).

The distribution of the observed data was the diameter of the inhibition zone from each variation of the incubation temperature and the overall data were normally distributed with a significance value of  $p$  (0.05). The data then continued with the homogeneity test, all of the data had sig values. (Mauchly's Test of Sphericity  $s$ ) 0.05 so that all data are homogeneous

**Table 3.** Statistical Test of Repeated ANOVA

<b>Antibiotic</b>	<b>Sig. (Sphericity Assumed)</b>
Clindamycin	0.010
Levofloxacin	0.000
Tetracycline	0.000
Trimethoprim	0.000

**Table 4.** Pairwise Comparison Test Results of temperature variations to an optimal temperature

<b>Antibiotic</b>	<b>Incubation Temperature</b>	<b>p-value</b>
Clindamycin	Inhibition zone diameter 33°C	0.137
	Inhibition zone diameter 34°C	0.162
	Inhibition zone diameter 36°C	0.668
	Inhibition zone diameter 37°C	0.219
Levofloxacin	Inhibition zone diameter 33°C	0.005
	Inhibition zone diameter 34°C	0.015
	Inhibition zone diameter 36°C	0.213
	Inhibition zone diameter 37°C	0.322
Tetracycline	Inhibition zone diameter 33°C	0.175
	Inhibition zone diameter 34°C	0.009
	Inhibition zone diameter 36°C	0.000
	Inhibition zone diameter 37°C	0.001
Trimethoprim	Inhibition zone diameter 33°C	0.001
	Inhibition zone diameter 34°C	0.010
	Inhibition zone diameter 36°C	0.735
	Inhibition zone diameter 37°C	0.780

Table 3 comparative test of Repeated ANOVA shows that sig. (Sphericity Assumed)  $<$  0.05, so it can be concluded that  $H_0$  is rejected. This means that there is a difference in the diameter of the inhibition zone of each antibiotic (Clindamycin, Levofloxacin, Tetracycline, and Trimethoprim) after incubation at 33°C, 34°C, 35°C, 36°C, and 37°C for 18 hours. While in table 4 it can be observed that the difference in the average variation of incubation temperature to the optimal temperature occurs in Levofloxacin antibiotics at 33°C and 34°C, in Tetracycline antibiotics at 34°C, 36°C, and 37°C, while on the antibiotic Trimethoprim, a significant difference was seen at 33°C and 34°C to the optimal temperature.

## DISCUSSION

Microbes can grow everywhere but are still influenced by environmental factors. Temperature is one of the environmental factors that affect microbial growth. Microorganisms can be controlled by an environment, by using various processes or physical means. Bacterial growth depends on chemical reactions and since the rate of these reactions is affected by temperature, bacterial growth is strongly influenced by temperature (Pelczar, 2008).

Every organism has the lowest temperature limit and the highest temperature, microbes need an optimum temperature for their growth and development. In general, bacteria grow at temperatures above 35°C, for each species, there is a maximum and minimum temperature limit for growth. The optimum temperature is closer to the maximum temperature while at the minimum temperature the growth is slower. If the temperature is higher than the maximum temperature, the growth of bacteria will decrease rapidly. This illustrates that temperature mainly affects enzymes, the higher the temperature, the faster the enzyme activity. If the temperature is too high then the enzyme will be denatured so that the cell will then die (Winarwi, 2006).

*Staphylococcus aureus* is a gram-positive bacterium with a diameter of 0.5-1.0 µm, in the form of a series of grapes, does not form spores, does not move, groups, pairs, and sometimes has short chains. Some strains of this bacterium have a capsule (Karimela et al., 2017). *Staphylococcus aureus* contains polysaccharide antigens and proteins as well as other substances that are important in the cell wall structure. Peptidoglycan, a polysaccharide polymer containing subunits joined together to form a rigid exoskeleton of the cell wall. Peptidoglycan can be damaged by strong acids or exposure to lysozyme. *Staphylococcus aureus* grows well on various bacteriological media under aerobic or microaerophilic conditions. Grows rapidly at 37°C, but the best pigment formation is at room temperature (20-35°C) (Riedel et al., 2019).

The four types of antibiotics used in this study, namely, Clindamycin, Levofloxacin, Tetracycline, and Trimethoprim are types of antibiotics that can be used to monitor the quality and purity of *Staphylococcus aureus* strains (Zimmer & et al, 2020). The selection of these four types of antibiotics is also based on the mechanism of action of antibiotics in inhibiting the growth of *Staphylococcus aureus* bacteria.

Clindamycin is the first lincomycin group of antibiotics to be introduced for clinical use, and its mechanism of action is very similar to that of erythromycin. Clindamycin belongs to a group of narrow-spectrum antibiotics, which are used for anaerobic and aerobic gram-positive bacteria. The mechanism of action of this antibiotic is to inhibit bacterial protein synthesis by reversibly binding to the 50S ribosomal subunit, thereby blocking the transpeptidation reaction or translocation of susceptible bacteria (Fanayoni et al., 2019).

Levofloxacin is a quinolone class of antibiotics that is useful for treating diseases caused by bacterial infections such as pneumonia, sinusitis, prostatitis, conjunctivitis, urinary tract infections, and skin infections. This fluoroquinolones class of antibiotics works by inhibiting the activity of bacterial topoisomerase (DNA gyrase) which is important in the process of bacterial replication (Toy, 2008).

Tetracycline is a broad-spectrum antibiotic used to treat many infectious diseases. This antibiotic works by inhibiting bacterial protein synthesis by inhibiting the binding of aminoacyl-tRNA to the bacterial 30s ribosome (Toy, 2008). Trimethoprim is a class of antibiotics that are bacteriostatic, but trimethoprim has a broad spectrum effect with activity against Gram-positive and Gram-negative organisms. This antibiotic works by binding to dihydrofolate reductase, a process that blocks the reduction of dihydrofolic acid to tetrahydrofolic acid. Interrupting the synthesis of tetrahydrofolic acid is important because these compounds play a central role in the pathway of thymidine synthesis and the synthesis of thymidine, which plays a role in bacterial DNA synthesis. In other words, by binding to dihydrofolate reductase, trimethoprim causes reduced bacterial folate synthesis (Toy, 2008).

The clear zone on the agar layer formed is caused by antimicrobial compounds diffusing into the agar layer and inhibiting the growth of microorganisms (bacteria) and is referred to as the inhibition zone, while the agar layer overgrown with microorganisms will appear cloudy. Antimicrobial compounds work by interacting with bacterial cell walls, resulting in impaired permeability of bacterial cell walls and facilitating antimicrobial compounds to diffuse into

bacterial cells. Diffusion that occurs will result in disruption of a series of growth processes of bacteria to inhibit their growth (bacteriostatic) or provide other effects, namely by killing bacteria (bacteriocidal). In addition, antimicrobial compounds can also penetrate cell membranes and interact with genetic material from bacteria so that bacteria can undergo mutations (Perdana & Setyawati, n.d.).

This study obtained the results that to obtain optimal growth, incubation must be carried out at a temperature of 35°C. Temperatures less than 35°C were found to be able to make most antibiotics form a wider diameter of the inhibition zone. This can be fatal in concluding and determining the choice of antibiotics that can be used for patient therapy. Resistance can be read susceptible if it is incubated at a low temperature, as a result, patient therapy is inadequate and the chances of developing antibiotic resistance are high (Nadjamuddin, 2023). Incubation of the media in the piles can result in the media in the middle of the dish not getting an even temperature distribution and the temperature of the media in the middle of the pile being low below 35°C. The higher the incubation temperature (> 35 ° C) the growth of germs can become fertile. The diameter of the inhibition zone formed is getting smaller, so it could be that antibiotics that should be used for patient therapy are reported to be resistant because the inhibition zone becomes smaller. In addition, high temperatures can also cause the diffusion of antibiotics into the Mueller-Hinton medium to be less good.

The bactericidal activities of ciprofloxacin and levofloxacin against *Staphylococcus aureus* and *epidermidis* were significantly reduced when incubated at lower temperatures. Levofloxacin was found to be more effective against both bacteria, suggesting it may be a more suitable option for staphylococcal infections, particularly at skin and soft-tissue sites (Parte & Smith, 1994). Other study found that isoniazid and rifampicin's bactericidal activities decreased at lower temperatures, while TMC207 was immediately bactericidal at 37°C. Pyrazinamide was more bactericidal at lower temperatures and dormant seed cultures. This suggests that low temperatures make bacteria more susceptible to TMC207's blocking of ATP synthesis and hindering the export of pyrazinoic acid (Coleman et al., 2011). Some research examined the impact of temperature on the adhesion and disinfection properties of Ag<sup>+</sup>-doped BiVO<sub>4</sub> coatings. The study found that the monoclinic scheelite phase remained unchanged after annealing at 450-650 °C. The silver-modified samples showed good disinfection activity, with the best adhesion and complete killing of *Escherichia coli*, *Staphylococcus aureus*, *Shigella*, and *Salmonella* after 2 hours of visible-light irradiation (Zhang et al., 2022).

The inhibition of bacterial growth by antibiotics can be seen from the clear zone formed around the antibiotic disc paper. The inhibition of the growth of bacterial colonies can be caused by damage to the structural components of the bacterial cell membrane. Damage to cell membranes disrupts nutrient transport (compounds and ions) so that bacterial cells experience a lack of nutrients needed for growth (Leboffe & Pierce, 2019). The time to kill *Escherichia coli* inoculum significantly increased after preservative dilution or temperature reduction, with a concentration exponent of 3 and a temperature coefficient of 2.3. (Lusher et al., 1984). The addition of sodium chloride and temperature significantly enhanced the bactericidal activity of spirit vinegar. The time required for a 3 log decrease in viable cell numbers was shortened by 5% sodium chloride and the rise of temperature. The study suggests that vinegar solution containing sodium chloride may be a useful method to prevent food poisoning by reducing the number of viable cells (Lusher et al., 1984).

The study shows that even at a low pH (4.4), ATSSB spores can grow without organic acids, but acetic and lactic acids have significant antibacterial activity. The study also found that a combination of these acids inhibits growth from *B. subtilis* and *B. velezensis*, making ambient storage of low-acid pasteurized sauces feasible (Sun et al., 2021). High temperature



aging of copper-bearing 2205 duplex stainless steel (Cu-2205 DSS) improved its corrosion resistance, mechanical properties, and antibacterial activity. Results showed increased  $\gamma$  phase in microstructure and formation of new  $\sigma$  phase and copper-rich precipitates. The mechanical properties were significantly enhanced, and aging increased uniform corrosion resistance but slightly affected pitting resistance. The antibacterial performance was improved due to increased release of copper ions, indicating potential applications in marine engineering (Khan et al., 2022). The antibacterial activity of SPR19, a strain of *Brevibacillus* sp., using atmospheric and room temperature plasma mutagenesis (ARTP). The 469 mutants were screened and confirmed, with M285 showing the highest activity. M285 was stable and tolerant to various conditions, retaining over 90% of its antibacterial properties. This suggests M285 as a potential antibacterial source (Songnaka et al., 2022). Another study reveals that microbial competition affects the synthesis of antimicrobial compounds by psychrophilic micromycetes, particularly in summer when higher temperatures increase total fungal activity. The most promising strain, *Penicillium vulpinum* KPB F-290, showed maximum antimicrobial activity against test cultures. It exhibited high antimicrobial activity against opportunistic strains and *B. subtilis* ATCC 6633, as well as phytopathogenic pectobacteria *P. carotovorum* and *P. savastanoi*. The optimal cultivation method for maximum antibiotic production is stationary, with ethyl acetate being the best extractant from the culture liquid (Kuvarina et al., 2022).

A multifunctional temperature-triggered antibacterial hemostatic fluorocopolymer aggregate coating was developed, consisting of fluoropolymer and quaternary ammonium salt. This coating has antibacterial, antibacterial, and hemostatic properties, promoting biocompatibility and anti-adhesion performance. It also has rapid coagulation, low blood loss, minor secondary bleeding, and minimal bacteria infiltration, making it potential for biomedical applications (Li et al., 2022).

In addition to the incubation temperature factor, several other factors can affect the formation of the diameter of the inhibition zone in the antibiotic sensitivity test, including the turbidity of the bacterial suspension, the time of absorption of the bacterial suspension into the MH media, the incubation time, the thickness of the MH media, the distance between the drug discs and the potency of the drug disc. In this study, control of other factors was carried out following standard procedures, such as the bacterial suspension was measured on a nephelometer with a turbidity of 0.5 McF, did not allow bacterial infiltration in MH media to exceed 15 minutes, the incubation time was not more than 18 hours, the thickness of the medium MH is only 4 mm or 20 ml of MH media is poured into the plate, the distance between the discs does not match because only four antibiotics are used which are always stored at  $-20^{\circ}\text{C}$  so as not to reduce the quality of the antibiotic potency (Hudzicki, 2009). Another research found the antimicrobial activity of the MgO-NPs increased significantly ( $p \leq 0.05$ ) with increased temperature, pH and NaCl concentration in TSB (Yoon et al., 2022).

The four types of antibiotics that have been selected are based on different mechanisms of action, besides based on the CLSI (2020) the four selected antibiotics namely Clindamycin, Levofloxacin, Tetracycline, and Trimethoprim can be used for Internal Quality Control of *Staphylococcus aureus* ATCC 25923. The yield range of the diameter of the QC inhibition zone for Clindamycin is 24-30 mm, Levofloxacin 25-30 mm, Tetracycline 24-30 mm, and Trimethoprim 19-26 mm (Zimmer & et all, 2020). The synthesized GONs demonstrated significant antibacterial activity against four bacterial strains using the agar well diffusion method. They also showed 60% higher radical scavenging activities (RSA) compared to Gallic acid, a standard used in the study (Goyat et al., 2022).

The results of the average diameter of the inhibition zone at temperatures of  $33^{\circ}\text{C}$ ,  $34^{\circ}\text{C}$ ,  $35^{\circ}\text{C}$ ,  $36^{\circ}\text{C}$ , and  $37^{\circ}\text{C}$  for all tested antibiotics were still within the allowable range, so the results of Quality Control of *Staphylococcus aureus* were acceptable. Appropriate antibiotic sensitivity testing is needed by clinicians to provide antibiotic treatment therapy to patients. To

guarantee sensitivity test results, routine QC testing using pure bacterial strains is very important in clinical microbiology. When selecting Quality Control strains and determining QC ranges, care is required for reproducibility of test results, the objective performance of using QC strains, strain stability, and so on (Hong et al., 2015).

The findings from this study demonstrate that incubation temperature significantly affects the diameter of the antibiotic inhibition zones in disc diffusion tests for *Staphylococcus aureus*. Specifically, the study observed that the inhibition zone diameters for Clindamycin, Levofloxacin, Tetracycline, and Trimethoprim varied with changes in temperature. For instance, the inhibition zone for Levofloxacin generally decreased as the incubation temperature increased, with significant reductions at 33°C and 34°C compared to the optimal temperature. Tetracycline showed a marked decrease in inhibition zone diameter at higher temperatures, particularly at 36°C and 37°C, indicating a strong temperature dependence. These temperature-related variations were statistically significant, as confirmed by repeated ANOVA tests ( $p < 0.05$  for all antibiotics), highlighting the importance of maintaining precise incubation conditions.

The high-temperature requirement A (HtrA) serine protease family is a promising target for antibacterial therapeutics due to their dual functions and diverse oligomerization patterns. A redesigned HtrA production method produces cleaner preparations with high yields by overexpressing and purifying target proteins under denaturing conditions. This method retains proteolytic and chaperone activity, allowing for higher production quantities and application in various protein purification strategies (Ronzetti et al., 2022).

The implications of these findings are profound for antibiotic sensitivity testing protocols. To ensure consistent and reliable results, laboratories must standardize incubation temperatures, as deviations can lead to significant differences in the measured effectiveness of antibiotics. This standardization is crucial for accurate diagnosis and appropriate treatment decisions in clinical settings, where inconsistent testing conditions could result in inappropriate antibiotic choices, potentially impacting patient outcomes. Furthermore, these results emphasize the need for rigorous quality control measures and updated guidelines in laboratories to monitor and regulate incubation temperatures meticulously. Ensuring that all personnel are trained on the importance of precise incubation conditions can reduce variability and improve the reliability of antibiotic sensitivity testing across different settings.

#### 4. CONCLUSION

There is a difference in the diameter of the antibiotic inhibition zone at various incubation temperatures. Based on these differences, it can be concluded that there is an effect of incubation temperature on the diameter of the antibiotic inhibition zone in the antibiotic sensitivity test of *Staphylococcus aureus* by the disc diffusion method. A high incubation temperature (above the optimal temperature) can make the average diameter of the antibiotic inhibition zone in the sensitivity test smaller, while a low incubation temperature (below the optimal temperature) makes the average diameter of the antibiotic inhibition zone in the sensitivity test wider. In addition, the temperature of  $\pm 2^\circ\text{C}$  from the optimal temperature of 35°C did not significantly affect the interpretation of the diameter of the antibiotic inhibition zone for the Internal Quality Control of *Staphylococcus aureus*. The research is limited by its focus on a single bacterial strain (*Staphylococcus aureus*) and a narrow range of antibiotics and temperatures, which may not fully represent the variability across different bacteria, antibiotics, and broader temperature ranges, potentially affecting the generalizability and robustness of the findings. Additionally, conducting the study in a single laboratory setting and maintaining a

constant incubation period without exploring the potential interaction with time also limits the comprehensiveness of the results.

## REFERENCES

- Coleman, D., Waddell, S. J., & Mitchison, D. A. (2011). Effects of low incubation temperatures on the bactericidal activity of anti-tuberculosis drugs. *Journal of Antimicrobial Chemotherapy*, 66(1), 146–150. <https://doi.org/10.1093/jac/dkq414>
- Fanayoni, A., Gelgel, K. T. P., & Suarjana, I. G. K. (2019). Uji Sensitivitas Bakteri *Staphylococcus* sp. asal Babi Penderita Porcine Respiratory Disease Complex terhadap Doxycycline, Kanamycin, dan Clindamycin. *Jurnal Indonesia Medicus Veterinus*, 8(4), 439-445.
- Goyat, R., Singh, J., Umar, A., Saharan, Y., Kumar, V., Algadi, H., Akbar, S., & Baskoutas, S. (2022). Modified low-temperature synthesis of graphene oxide nanosheets: Enhanced adsorption, antibacterial and antioxidant properties. *Environmental Research*, 215, 114245. <https://doi.org/10.1016/j.envres.2022.114245>
- Hong, S. K., Choi, S. J., Shin, S., Lee, W., Pinto, N., Shin, N., Lee, K., Hong, S. G., Kim, Y. A., Lee, H., Kim, H., Song, W., Lee, S. H., Yong, D., Lee, K., & Chong, Y. (2015). Establishing Quality Control Ranges for Antimicrobial Susceptibility Testing of *Escherichia coli*, *Pseudomonas aeruginosa*, and *Staphylococcus aureus*: A Cornerstone to Develop Reference Strains for Korean Clinical Microbiology Laboratories. *Annals of Laboratory Medicine*, 35(6), 635–638. <https://doi.org/10.3343/alm.2015.35.6.635>
- Hudzicki, J. (2009). Kirby-Bauer disk diffusion susceptibility test protocol. *American Society for Microbiology*, 15(1), 1–23.
- Jenul, C., & Horswill, A. R. (2019). Regulation of *Staphylococcus aureus* virulence. *Microbiology spectrum*, 7(2), 10.1128. <https://doi.org/10.1128/microbiolspec.GPP3-0031-2018>
- Karimela, E. J., Ijong, F. G., & Dien, H. A. (2017). Characteristics of *Staphylococcus aureus* Isolated Smoked Fish Pinekuhe from Traditionally Processed from Sangihe District. *Jurnal Pengolahan Hasil Perikanan Indonesia*, 20(1), 188. <https://doi.org/10.17844/jphpi.v20i1.16506>
- Khan, M. S., Yang, C., Pan, H., Yang, K., & Zhao, Y. (2022). The effect of high temperature aging on the corrosion resistance, mechanical property and antibacterial activity of Cu-2205 DSS. *Colloids and Surfaces B: Biointerfaces*, 211, 112309. <https://doi.org/10.1016/j.colsurfb.2021.112309>
- Kuvarina, A. E., Roshka, Yu. A., Rogozhin, E. A., Nikitin, D. A., Kurakov, A. V., & Sadykova, V. S. (2022). Antimicrobial Properties and the Effect of Temperature on the Formation of Secondary Metabolites in Psychrophilic Micromycetes. *Applied Biochemistry and Microbiology*, 58(3), 243–250. <https://doi.org/10.1134/S0003683822030085>
- Leboffe, M. J., & Pierce, B. E. (2019). *Microbiology: Laboratory Theory and Application, Essentials*. Morton Publishing Company. <https://books.google.co.id/books?id=8ZOFDwAAQBAJ>
- Lenggu, C. K. L., Indriarini, D., & Amat, A. L. S. (2020). Uji Aktivitas Ekstrak Etanol Kulit Daging Buah Lontar (*Borassus Flabellifer* Linn) Terhadap Pertumbuhan *Escherichia Coli* Secarain Virto. *Cendana Medical Journal*, 8(2), 96-107. <https://doi.org/doi.org/10.35508/cmj.v8i2.3353>
- Li, W., Zhang, Y., Ding, J., Zhang, S., Hu, T., Li, S., An, X., Ren, Y., Fu, Q., Jiang, X., & Li, X. (2022). Temperature-triggered fluorocopolymer aggregate coating switching from antibacterial to antifouling and superhydrophobic hemostasis. *Colloids and Surfaces B: Biointerfaces*, 215, 112496. <https://doi.org/10.1016/j.colsurfb.2022.112496>

- Lusher, P., Denyer, S. P., & Hugo, W. B. (1984). A note on the effect of dilution and temperature on the bactericidal activity of potassium sorbate. *Journal of Applied Bacteriology*, 57(1), 179–181. <https://doi.org/10.1111/j.1365-2672.1984.tb02372.x>
- Murray, P. R., Rosenthal, K. S., & Pfaller, M. A. (2021). *Medical microbiology* (Ninth edition). Elsevier.
- Nadjamuddin, M. (2023). *Pengantar Bakteriologi* (1st ed.). Eureka Media Aksara.
- Notoatmodjo, S. (2005). *Metodologi penelitian kesehatan*. Jakarta: Rineka Cipta.
- Novard, M. F. A., Suharti, N., & Rasyid, R. (2019). Gambaran Bakteri Penyebab Infeksi Pada Anak Berdasarkan Jenis Spesimen dan Pola Resistensinya di Laboratorium RSUP Dr. M. Djamil Padang Tahun 2014–2016. *Jurnal Kesehatan Andalas*, 8(2S), 26. <https://doi.org/10.25077/jka.v8i2S.955>
- Parte, A. C., & Smith, J. T. (1994). Effects of temperature on the bactericidal activities of ciprofloxacin and levofloxacin against *Staphylococcus aureus* and *Staphylococcus epidermidis*. *Microbios*, 79(319), 87–95.
- Pelczar, M. J. (2008). *Dasar-dasar mikrobiologi*. Penerbit Universitas Indonesia.
- Perdana, R., & Setyawati, T. (n.d.). Uji in-vitro sensitivitas antibiotik terhadap bakteri *Salmonella Typhi* di Kota Palu. *Medika Tadulako: Jurnal Ilmiah Kedokteran Fakultas Kedokteran Dan Ilmu Kesehatan*, 3(1), 11–22.
- Riedel, S., Hobden, J. A., Miller, S., Morse, S. A., Mietzner, T. A., Jawetz, E., Melnick, J. L., & Adelberg, E. A. (Eds.). (2019). *Jawetz, Melnick & Adelberg's Medical Microbiology* (28th edition). McGraw Hill.
- Ronzetti, M., Baljinnnyam, B., Jalal, I., Pal, U., & Simeonov, A. (2022). Application of biophysical methods for improved protein production and characterization: A case study on an high-temperature requirement A-family bacterial protease. *Protein Science*, 31(12), e4498. <https://doi.org/10.1002/pro.4498>
- Siregar, M. T., Wulan, W. S., Setiawan, D., & Nuryati, An. (2018). Kendali Mutu. *Buku Ajar Teknologi Laboratorium Medik (TLM)*, Kemenkes RI, Jakarta.
- Smith, P., Finnegan, W., Ngo, T., & Kronvall, G. (2018). Influence of incubation temperature and time on the precision of MIC and disc diffusion antimicrobial susceptibility test data. *Aquaculture*, 490, 19–24. <https://doi.org/10.1016/j.aquaculture.2018.02.020>
- Songnaka, N., Nisoa, M., Atipairin, A., Wanganuttara, T., & Chinnawong, T. (2022). Enhanced Antibacterial Activity of *Brevibacillus* sp. SPR19 by Atmospheric and Room Temperature Plasma Mutagenesis (ARTP). *Scientia Pharmaceutica*, 90(2), 23. <https://doi.org/10.3390/scipharm90020023>
- Sun, R., Vermeulen, A., & Devlieghere, F. (2021). Modeling the combined effect of temperature, pH, acetic and lactic acid concentrations on the growth/no growth interface of acid-tolerant *Bacillus* spores. *International Journal of Food Microbiology*, 360, 109419. <https://doi.org/10.1016/j.ijfoodmicro.2021.109419>
- Toy, E. C. (Ed.). (2008). *Case files: Microbiology: [50 clinical cases with USMLE-style questions help you ace course exams and the boards, microbiology pearls highlight key points, primer teaches you how to approach clinical problems, proven learning system maximizes shelf exam scores]* (2nd ed). McGraw-Hill Medical.
- Wiley, J. M., Sherwood, L., & Woolverton, C. J. (2017). *Prescott's microbiology* (Tenth edition). McGraw-Hill.
- Winarwi, W. (2006). *Uji Viabilitas Bakteri dan Aktivitas Enzim Bakteri Proteolitik pada Media Carrier Bekatul (sebagai Acuan Bahan Ajar Pokok Bahasan Virus, Monera, Protista di SMA)* [Skripsi]. Universitas Sebelas Maret.

- Yoon, J. Y., Yeom, W., Kim, H., Beuchat, L. R., & Ryu, J.-H. (2022). Effects of temperature, pH and sodium chloride on antimicrobial activity of magnesium oxide nanoparticles against *E. coli* O157:H7. *Journal of Applied Microbiology*, 133(4), 2474–2483. <https://doi.org/10.1111/jam.15719>
- Zhang, Y., Zhao, X., Wang, H., Fu, S., Lv, X., He, Q., Liu, R., Ji, F., & Xu, X. (2022). Effect of Temperature on the Adhesion and Bactericidal Activities of Ag<sup>+</sup>-Doped BiVO<sub>4</sub> Ceramic Tiles. *Inorganics*, 10(5), 61. <https://doi.org/10.3390/inorganics10050061>
- Zimmer, B., & et all. (2020). *Performance Standards for Antimicrobial Disk Susceptibility Tests* (12th ed.). Clinical and Laboratory Standards Institute.